

## TECHNICAL DATASHEET

### 004546

Material:	Polyethylen / laminate film, blue
Others:	Size 2XL
Packaging unit:	50 pieces



Type PB [6]-B



Shell fabric: 100% polyethylene

### Description

NITRAS medical gown, blue (colour code: 2200), polyethylene, with laminat film, pleasant wearing comfort, breathable material, round collar, cuffs made of soft cotton, to tie at the neck and behind the body, length approx. 110 cm, individually packed in bags, protection against infection agents, type PB [6]-B

	Material	Colour
Shell fabric:	100% polyethylene	Blue



Do not wash



Do not bleach



Do not tumble dry



Do not iron



Do not dry clean

#### Kreissparkasse Köln

IBAN: DE49 3705 0299 0127 2721 64  
BIC: COKSDE33XXX

#### Sparkasse Köln/Bonn

IBAN: DE32 3705 0198 1933 3245 90  
BIC: COLSDE33XXX

#### HypoVereinsbank

IBAN: DE63 3702 0090 0027 7497 39  
BIC: HYVEDEMM429

Es gelten unsere allgemeinen Geschäfts- und Lieferbedingungen (zu finden unter [www.nitras.de](http://www.nitras.de) oder auf Anfrage). Our general terms and conditions of business and delivery apply (can be found at [www.nitras.de](http://www.nitras.de) or on request).


MCD Medical Care Dental GmbH | Heinrich-Hertz-Str. 7 | 50181 Bedburg | Germany | Geschäftsführer / Managing Director: Maria Baric, Markus Seiler | HRB 83971 | Gerichtsstand Köln | USt.-IdNr.: DE299692222


+49 2272 401 111

[info@nitras-medical.com](mailto:info@nitras-medical.com)

[www.nitras-medical.com](http://www.nitras-medical.com)

**EN ISO 13688:2013** Protective clothing - General requirements

<b>EN 13034:2005 + A1:2009</b>	Test parameter	Performance level	Test result
 <p><b>Type PB [6]-B</b></p>	Abrasion resistance	1-6	3
	Tear resistance (trapezoidal)	1-6	2
	Tensile strength	1-6	1
	Puncture resistance	1-6	1
	Flex cracking resistance	1-6	6
	Seams strength	1-6	1

<b>EN 14126:2003</b>	Test parameter	Performance level	Test result
	Resistance to penetration by contaminated liquids: Synthetic blood	1-6	6
	Resistance to penetration by contaminated liquids: Phi-X174 bacteriophage	1-6	6
	Resistance to penetration by infective agents due to mechanical contact with substances containing contaminated liquids: Staphylococcus aureus	1-6	6
	Resistance to penetration by contaminated aerosols: Staphylococcus aureus	1-3	3
	Resistance to penetration by contaminated solid particles: Bacillus subtilis	1-3	3

**Kreissparkasse Köln**

IBAN: DE49 3705 0299 0127 2721 64  
BIC: COKSDE33XXX

**Sparkasse Köln/Bonn**

IBAN: DE32 3705 0198 1933 3245 90  
BIC: COLSDE33XXX

**HypoVereinsbank**

IBAN: DE63 3702 0090 0027 7497 39  
BIC: HYVEDEMM429

Es gelten unsere allgemeinen Geschäfts- und Lieferbedingungen (zu finden unter [www.nitras.de](http://www.nitras.de) oder auf Anfrage). Our general terms and conditions of business and delivery apply (can be found at [www.nitras.de](http://www.nitras.de) or on request).